



# Facility Requirements

	A220	A440	A700	A1000
Space Required - Steps Up: L x W x H - in (mm)	34.0 x 36.0 x 78.0 (864 x 914 x 1956)	45.0 x 47.0 x 79.0 (1143 x 1194 x 2007)	55.0 x 62.5 x 80.0 (1397 x 1588 x 2032)	65.0 x 72.2 x 81.0 (1651 x 1834 x 2057)
Space Required - Steps Down: L x W x H - in (mm)	34.0 x 44.5 x 78.0 (864 x 1130 x 1956)	45.0 x 55.0 x 79.0 (1143 x 1397 x 2007)	55.0 x 74.0 x 80.0 (1397 x 1880 x 2032)	65.0 x 83.7 x 81.0 (1651 x 2126 x 2057)
Max Floor Load lbs/ft <sup>2</sup> (kg/m <sup>2</sup> )	378 (1846)	352 (1719)	322 (1572)	305 (1490)
	E264	E528	E840	E1200
Space Required - Steps Up: L x W x H - in (mm)	34.0 x 40.0 x 82.0 (864 x 1016 x 2083)	45.0 x 52.0 x 83.0 (1143 x 1321 x 2108)	55.0 x 62.5 x 84.0 (1397 x 1588 x 2134)	65.0 x 72.2 x 85.0 (1651 x 1834 x 2159)
Space Required - Steps Down: L x W x H - in (mm)	34.0 x 50.0 x 82.0 (864 x 1270 x 2083)	45.0 x 62.5 x 83.0 (1143 x 1588 x 2108)	55.0 x 74.0 x 84.0 (1397 x 1880 x 2134)	65.0 x 83.7 x 85.0 (1651 x 2126 x 2159)
Max Floor Load lbs/ft <sup>2</sup> (kg/m <sup>2</sup> )	454 (2217)	422 (2060)	386 (1885)	366 (1787)
Electrical	100-230VAC   50-60Hz   0.7A continuous, 2.9A full load			
LN2 Supply	22-35 PSI (1.5-2.4 Bar)   CGA-295			
Ventilation	Laboratory air change rate minimums are typically sufficient for LN2 freezer installations; however, consult your Health and Human Safety Officer or equivalent.			
Oxygen Monitoring	Oxygen monitoring and alarms are recommended for any LN2 setup.			
Temperature	Near room temperature, 18°-27°C (65°-80°F)			
Relative Humidity	Less than 50%, non condensing			
Seismic Restraint	Seismic restraints may be required in earthquake prone areas. Please follow local standards and regulations. Contact Azenta for more information.			

## Regulatory

Azenta Cryo Freezers meet the following safety requirements conforming to:  
UL STD 61010-1, CSA STD C22.2 # 61010-1, Low Voltage Directive (2014/35/EU)